

HD wheelchairs

Adaptation and Restoration

Art. No. 95109-1

	<p>General instructions for adapting, repairing and restoring CE marked HD wheelchairs.</p>	
	<p>Adaptation</p>	<p>See document 95102-1. The customer/the customer organisation is responsible for ensuring that personnel have appropriate competence to carry out the adaptations.</p>
	<p>Repairs</p>	<p>The customer/the customer organisation is responsible for ensuring that personnel have appropriate competence to carry out the repairs. Personnel should have undergone Hantverksdesign's one-day course for technicians. Material and components: Hantverksdesign is only responsible for spare parts and components that we have supplied. Normal machine components such as screws, nuts and washers can be purchased from other suppliers if the names and numbers in the Spare Parts List are complied with. The latest revisions of the relevant documents must be complied with.</p>
	<p>Restoration</p>	<p>The following is required to restore specially adapted HD wheelchairs to become CE marked wheelchairs. Authorisation agreements must have been entered into. The customer/the customer organisation is responsible for ensuring that personnel have appropriate competence to restore the wheelchair. Personnel must have undergone Hantverksdesign's one-day course for technicians. The restoration work must be approved by Hantverksdesign. Material and components: Only original parts may be used for restoring wheelchairs to become CE-marked wheelchairs. Restoration means that the wheelchair is restored to its previous standard specifications, or is updated to a later version's standard specifications.</p>

Hantverksdesign & Rehab.prod. AB

Box 1263, SE-181 24 Lidingö, Sweden
info@hantverksdesign.se

Tel: +46 (0)8-767 04 80 Fax: +46 (0)8-767 50 00
www.hantverksdesign.se